

# **IV Stands Care & Cleaning Instructions**

#### **IV Stand Series**

Stainless Steel

CuVerro®

- P-1080
- P-1090
- P-1572
- P-1576

P-1083

### Caution

**DO NOT CART WASH**. Product may be damaged by exposure to the high temperatures.

# **Disinfection**

Disinfection may be achieved by use of hospital grade chemical disinfectants. This includes up to 10% bleach solution or bleach wipes. Refer to the disinfectant manufacturer's instructions for use.

## **Equipment Required:**

Disinfectant, water dampened cloth

#### Disinfectant Instructions:

- 1. Apply the disinfectant to the product per manufacturer's instructions.
- 2. Let the product stand for the exposure time per the manufacturer's instructions.
- 3. Wipe the product with a water dampened cloth to remove disinfectant residual.

  (It is recommended that the product be wiped down using a water dampened cloth after disinfection to prevent possible corrosion of the product's materials caused by overexposure to the disinfectant.)

# **Cleaning**

Cleaning is required only if the product is heavily soiled and standard disinfection is not sufficient. Be sure to follow the detergent manufacturer's instructions for use.

## **Equipment Required:**

Detergent, soft bristle brush, lint free cloth, running water

# Cleaning Instructions:

- 1. Rinse excess soil from all surfaces
- 2. Apply detergent per manufacturer's instructions. Apply to all surfaces and use the brush to scrub surfaces as required. (Brushing along the material grain will reduce the noticeability of any scratches.)
- 3. Thoroughly rinse all surfaces of residual detergent and soil. See manufacturer's instructions for details.
- 4. Dry with lint free cloth or per manufacturer's instructions.

#### CuVerro® IV Stand Series\*

• P-1083

\*Accessories produced with CuVerro® in place of standard materials: P-1083-C, P-1083-F, P-3303

Disinfecting and cleaning instructions for CuVerro® materials are the same as the above stainless steel products, but a citrus-based cleaner is recommended for best results.

(CleanCide, manufactured by Wexford Labs, Inc. is recommended. CleanCide disinfects and also protects the CuVerro® surface, minimizing the discoloration that may occur over time.)

PN: 0957-00251-00